



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ATLAS MATERIAL TESTING TECHNOLOGY, LLC
1500 Bishop Court
Mount Prospect, IL 60056
Peter Wysgalla Phone: 773 289 5720

CALIBRATION

Valid To: November 30, 2023

Certificate Number: 2101.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 8}:

I. Optical Radiation

Parameter/Equipment	Range	CMC ^{2, 7} (±)	Comments ⁶
Control Parameters in Weathering Instruments ^{3, 4} –			
Temperature	(0 to 85) °C	0.8 °C	Fluke 51 digital thermometer
Relative Humidity	(5 to 90) % RH	3.9 % RH	Vaisala HMI 51 humidity calibrator
AC Power ⁵	Up to 6 kW (6 to 12) kW	0.4 % 3.0 %	Hioki PW3335 power meter
Pressure	Up to 20 psig	0.6 psig	Omega DPG1000B-100G digital pressure gauge

Parameter/Equipment	Range	CMC ^{2,7} (±)	Comments ⁶
Illuminance ^{3,4} (with Xenocal) – SUNTEST Instruments (380 to 780) nm Xenotest Instruments (380 to 780) nm	Up to 240 klx Up to 240 klx	5.0 % 7.6 % 10 % 4.6 %	Daylight, SolarStandard, WG, StoreLight, ID65 XC270, XC300, B04 XC320, XC320HLF, 7IR, 16H, 4IR3WG, 6IR1UV, GMW3414 10WG, TM16, DL Ext. IR
Irradiance ^{3,4} (with Xenocal) – SUNTEST Instruments 340 nm 420 nm (300 to 400) nm (300 to 800) nm (300 to 400) nm (300 to 800) nm Xenotest Instruments – 340 nm 420 nm (300 to 400) nm (300 to 800) nm (300 to 400) nm (300 to 800) nm (300 to 400) nm (300 to 800) nm (300 to 400) nm (300 to 800) nm	Up to 1.4 W·m ⁻² ·nm ⁻¹ Up to 3.5 W·m ⁻² ·nm ⁻¹ Up to 150 W·m ⁻² Up to 1300 W·m ⁻² Up to 150 W·m ⁻² Up to 1300 W·m ⁻² Up to 1.4 W·m ⁻² ·nm ⁻¹ Up to 3.5 W·m ⁻² ·nm ⁻¹ Up to 150 W·m ⁻² Up to 1300 W·m ⁻² Up to 150 W·m ⁻² Up to 1300 W·m ⁻² Up to 150 W·m ⁻² Up to 1300 W·m ⁻²	8.1 % 7.5 % 11 % 8.0 % 14 % 8.0 % 8.1 % 7.5 % 14 % 12 % 9.9 % 9.8 % 11 % 7.8 % 9.3 % 7.8 %	Daylight, SolarStandard, WG, StoreLight ID65 XC270, XC300, XC320, XC320HLF, 7IR, 16H 4IR3WG, 6IR1UV, GMW3414, 10WG, TM16, DL Ex. IR, B04 XC320, XC320HLF, 7IR, 16H, 4IR3WG, 6IRiUV, GMW3414 XC270, XC300, B04 10WG, TM16 DL Ext. IR

Parameter/Equipment	Range	CMC ^{2,7} (±)	Comments ⁶
Irradiance ^{3,4} (with Xenocal) – (cont) Ci Instruments 340 nm 420 nm (300 to 400) nm	Up to 1.4 W·m ⁻² ·nm ⁻¹ Up to 3.5 W·m ⁻² ·nm ⁻¹ Up to 150 W·m ⁻²	8.9 % 8.0 % 11 %	Boro S/Boro-S
Irradiance ^{3,4} (with Reference Lamp) – 340 nm 420 nm (300 to 400) nm	Up to 3.3 W·m ⁻² ·nm ⁻¹ Up to 7.8 W·m ⁻² ·nm ⁻¹ Up to 400 W·m ⁻²	5.8 % 5.5 % 5.6 %	Hioki PW3335 wattmeter w/Hioki 9660 current probe Boro S/Boro-S
Irradiance ^{3,4} (with Reference Radiometer) – Fluorescent Instruments (310, 340, 351) nm 254 nm	Up to 3.0 W·m ⁻² ·nm ⁻¹ Up to 3.0 W·m ⁻² ·nm ⁻¹ Up to 16.0 mW·cm ⁻² ·nm ⁻¹	7.6 % 13 % 9.7 %	UV Test UV2000 UVC Test

Parameter/Equipment	Range	CMC ^{2, 7} (±)	Comments ⁶
Irradiance (Calibration Standards) –			
Customer Xenon Ref. Lamps for Ci Instruments Operating at –			
Lamp AC Power up to 6 kW			
340 nm	Up to 3.3 W·m ⁻² ·nm ⁻¹	5.5 %	SP320 instrument systems spectroradiometer, Hioki PW3335 wattmeter w/Hioki 9660 current probe & 2 working standards Boro S/Boro-S
420 nm	Up to 7.8 W·m ⁻² ·nm ⁻¹	4.4 %	
(300 to 400) nm	Up to 400 W·m ⁻²	4.9 %	
Lamp AC Power up to 12 kW			
340 nm	Up to 3.3 W·m ⁻² ·nm ⁻¹	5.8 %	
420 nm	Up to 7.8 W·m ⁻² ·nm ⁻¹	5.5 %	
(300 to 400) nm	Up to 400 W·m ⁻²	5.6 %	
Customer Ref. UV Radiometers for Fluorescent Instruments –			
UV Test Fluorescent Instrument			
310/340/351 nm	Up to 3.0 W·m ⁻² ·nm ⁻¹	7.6 %	SP320 instrument systems, spectroradiometer and 3 working standards
UV2000 Fluorescent Instrument			
310/340/351 nm	Up to 3.0 W·m ⁻² ·nm ⁻¹	13 %	
UVC Test Instrument	Up to 16.0 mW·cm ⁻² ·nm ⁻¹	9.7 %	ILT2400 photometer and 3 working standards

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC Uncertainty due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the Calibration and Measurement Capability Uncertainty (CMC) found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC. Atlas Global Technical Service team members, residing in the following countries, fall under the

Atlas ISO17025 accreditation scope: Germany, France, Switzerland, Austria, Netherlands, United Kingdom, India, and the United States.

⁴ This includes but is not limited to all Atlas Weather-Ometer® and Fade-Ometer® instruments, UVTest, UV2000, Xenotest, and SUNTEST instruments.

⁵ AC Power are calibrated in the artificial weathering equipment to control temperature, humidity and irradiance.

⁶ Methods of calibration include the use of equipment listed in the column or equivalent.

⁷ In the statement of CMC Uncertainty, all percentages are defined as “percent of reading”.

⁸ This scope meets A2LA’s *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

ATLAS MATERIAL TESTING TECHNOLOGY LLC

Mount Prospect, IL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets *R205 – Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 10th day of December 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2101.01
Valid to November 30, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.