



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

ATLAST MATERIAL TESTING TECHNOLOGY LLC  
ATLAS WEATHERING SERVICES GROUP  
45601 North 47<sup>th</sup> Avenue  
Phoenix, AZ 85087-7042  
Lorenzo Tyler Phone: 305 245 3659

CALIBRATION

Valid To: September 30, 2020

Certificate Number: 0717.04

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,3,7</sup>:

I. Electrical DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2,4,5</sup> (±)	Comments
DC Voltage – Generate	Up to 202 mV (0.2 to 2.02) V (2 to 20.2) V (20 to 202) V (200 to 1025) V	0.003 % 0.004 % 0.002 % 0.002 % 0.002 %	Transmille 3010A
DC Current – Generate	Up to 202 uA (0.2 to 2.02) mA (2 to 20.2) mA (20 to 202) mA (0.2 to 2.02) A (2 to 10) A	0.3 % 0.16 % 0.48 % 0.13 % 0.03 % 0.04 %	Transmille 3010A

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> (±)	Comments
AC Voltage – Generate			
(0 to 202) mV	(10 to 44) Hz (1 to 20) kHz (20 to 100) kHz (100 to 500) kHz	0.21 % 0.08 % 0.29 % 2.5 %	Transmille 3010A
(0.2 to 2.0) V	(10 to 44) Hz (1 to 20) kHz (20 to 100) kHz (100 to 500) kHz	0.08 % 0.11 % 0.11 % 0.34 %	
(2 to 20.2) V	(10 to 44) Hz 45 Hz to 1 kHz (20 to 100) kHz	0.40 % 0.10 % 0.11 %	
(20 to 202) V	(10 to 44) Hz (1 to 20) kHz (20 to 100) kHz	0.04 % 0.06 % 0.17 %	
(200 to 750) V	(30 to 44) Hz 45 Hz to 1 kHz (1 to 10) kHz	0.03 % 0.05 % 0.05 %	
AC Current – Generate			
(2 to 20.2) mA	(10 to 44) Hz (1 to 10) kHz	0.43 % 0.32 %	Transmille 3010A
(20 to 202) mA	(10 to 44) Hz (1 to 10) kHz	0.41 % 0.54 %	
(0.2 to 2.02) A	(10 to 45) Hz (1 to 10) kHz	0.08 % 0.52 %	
(2 to 30) A	(30 to 45) Hz 100 Hz to 1 kHz	0.09 % 0.51 %	
Capacitance – Generate			
1 nF 10 nF	1 kHz	0.33 % 0.12 %	Transmille 3010A

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> (±)	Comments
Resistance – Generate	100 Ω 1 kΩ 10 kΩ 100 kΩ	0.01 % 0.001 % 0.024 % 0.34 %	Transmille 3010A
2 Wire Only	1 MΩ 10 MΩ	0.001 % 0.20 %	
DC Voltage – Measure	100 mV to 1000 V	0.002 %	Fluke 8845A
Resistance – Measure	100 Ω to 100 MΩ	0.01 %	Fluke 8845A
DC Current – Measure	(-100 to 0) μA (0 to 100) μA 100 μA to 10 A	0.20 % 0.13 % 0.48 %	Fluke 8845A

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> (±)	Comments
AC Current – Measure			
10 mA to 10 A	10 Hz to 10 kHz	0.55 %	Fluke 8845A
AC Voltage – Measure			
100 mV to 750 V	10 Hz to 20 kHz (20 to 50) kHz (100 to 300) kHz	0.11 % 0.29 % 0.66 %	Fluke 8845A

II. Optical Quantities

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
<p>Direct Solar Irradiance – Measure</p> <p>Secondary Pyrheliometers</p> <p>Field Pyrheliometers</p>	<p>(0.29 to 3.0) μm</p> <p>(0.29 to 3.0) μm</p>	<p>0.67 %</p> <p>0.88 %</p>	<p>ASTM E816; ISO 9059</p> <p>Eppley AHF primary cavity pyrheliometer</p> <p>Basis: 1000 W/m<sup>2</sup> day</p> <p>Secondary pyrheliometer</p> <p>Basis: 1000 W/m<sup>2</sup> day</p>
Field Pyranometers	<p>(0.29 to 3.0) μm</p> <p>Normal Incidence (31 to 45)° Tilt</p>	<p>1.1 %</p> <p>1.6 %</p>	<p>ASTM E824; ISO 9847</p> <p>Group of reference pyranometers (two or more)</p> <p>Basis: 1000 W/m<sup>2</sup> day</p>
<p>Global Solar Irradiance – Measure</p> <p>Reference Pyranometers:</p>	<p>(0.29 to 3.0) μm</p> <p>Normal Incidence (0 to 45)° Tilt</p>	<p>0.78 %</p> <p>1.1 %</p>	<p>ASTM G167; ISO 9846</p> <p>Eppley AHF primary cavity pyrheliometer</p> <p>Basis: 1000 W/m<sup>2</sup> day</p>
<p>Spectral Irradiance – Measure</p> <p>Spectroradiometers (Normal Incidence to Source)</p> <p>Solar Simulators, Incl, Classification</p>	<p>(0.280 to 0.315) μm</p> <p>(0.280 to 0.400) μm</p> <p>(0.295 to 0.385) μm</p> <p>(0.300 to 0.400) μm</p> <p>(0.315 to 0.400) μm</p> <p>(0.250 to 0.800) μm</p> <p>(0.380 to 1.100) μm</p> <p>(0.250 to 0.800) μm</p> <p>(0.380 to 1.100) μm</p>	<p>3.6 %</p> <p>3.6 %</p> <p>3.6 %</p> <p>3.5 %</p> <p>3.5 %</p> <p>3.7 %</p> <p>3.6 %</p> <p>3.7 %</p> <p>3.6 %</p>	<p>ASTM G138</p> <p>Standard of spectral irradiance (lamp), OL-752-10E</p> <p>ASTM E927; IEC 60904-9 OL 754 &amp; OL 770 Spectroradiometers</p>

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
Spectral Irradiance – Measure (cont)			
Reference Ultra-Violet Radiometers	(0.280 to 0.315) μm (0.280 to 0.400) μm (0.295 to 0.385) μm (0.300 to 0.400) μm (0.315 to 0.400) μm	4.2 % 3.7 % 3.7 % 3.6 % 3.9 %	ASTM G 130 OL 754 Spectroradiometer Basis: 0.2 W/m <sup>2</sup> Basis: 70 W/m <sup>2</sup> Basis: 60 W/m <sup>2</sup> Basis: 70 W/m <sup>2</sup> Basis: 70 W/m <sup>2</sup>
Field Ultra-Violet Radiometer –			ASTM E824; ISO 9847 (Kipp & Zonen UV-S-AB-T & UV-S-A-T (or other) reference UV radiometers
UV-B Ultra Violet Radiometers	(0.280 to 0.315) μm	4.2 %	Basis: 0.2 W/m <sup>2</sup>
Total Ultra Violet Radiometers	(0.295 to 0.385) μm	3.7 %	Basis: 60 W/m <sup>2</sup>
UV-A Ultra Violet Radiometers	(0.300 to 0.400) μm (0.315 to 0.400) μm	3.6 % 3.9 %	Basis: 70 W/m <sup>2</sup> Basis: 70 W/m <sup>2</sup>

### III. Time & Frequency

Parameter/Equipment	Frequency	CMC <sup>2, 4, 6</sup> (±)	Comments
Frequency – Generate			
Sine Wave	1 Hz to 1 mHz	0.0004 %	HP 33120A
Square Wave	1 Hz to 1 mHz	0.0007 %	

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Calibrations are performed under reasonably controlled conditions of clear skies to minimize atmospheric effects, high solar radiance approaching 1000 W/m<sup>2</sup> to preclude linearity effects and near-normal incidence for all calibrations unless stated otherwise to preclude cosine effects. It should be recognized that the uncertainties associated with the use of carefully calibrated radiometers under the diverse conditions of field use would be significantly higher.

<sup>4</sup> In the statement of CMC, percentages are to be read as a percentage of the reading unless otherwise noted.

<sup>5</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

<sup>6</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

<sup>7</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## Accredited Laboratory

A2LA has accredited

### ATLAS MATERIAL TESTING TECHNOLOGY LLC ATLAS WEATHERING SERVICES GROUP

*Phoenix, AZ*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 28<sup>th</sup> day of June 2018.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 0717.04  
Valid to September 30, 2020  
Revised September 08, 2020

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*